

Reference No. : WTF20F11091760J
Applicant : Foshan MyCarol Electric Technologies Inc.
Address : Room501&502, Block Two, Qide Industrial Park,
No.2 of Xingye East Road, Beijiao Town, Shunde
District Foshan City, Guangdong P.R. China
Manufacturer : The same as above
Address : The same as above
Product : Electric Breast Pump
Model No. : XN-D250, XN-D251, XN-D252, XN-D253, XN-D254,
XN-D255, XN-D256, XN-D257, XN-D258, XN-D259
Technical data : For adapter: Input: 100-240V~, 50/60Hz, 0.25A;
Output: DC 5V, 1A, 5W; Class II
For Electric Breast Pump: Input: DC 5V, 1A

Test Standards:

EN 60335-1:2012+A11:2014+A13:2017+A1:2019+A14:2019+A2:2019
EN 62233:2008

The above product has been tested by us with the listed standards and found in compliance with the European Directive LVD 2014/35/EU. It is possible to use CE marking to demonstrate the compliance with this LVD Directive.

Household and similar electrical appliances**Part 1: General requirements****Measurement methods for electromagnetic fields of household appliances and similar apparatus with regard to human exposure**

The referred test report(s) show that the product complies with standard(s) recognized as giving presumption of compliance with the essential requirements in the above mentioned EU Directive. Other relevant Directives have to be observed.

After preparation of the necessary technical documentation as well as the EU Declaration of Conformity, the CE marking as shown below can be affixed on the equipment under the sole responsibility of the manufacturer.



Manager: **Jerry Mu**
Date: Mar. 09, 2021

The statement is based on a single evaluation of the sample of above mentioned product. It does not imply an assessment of the whole production.

Waltek Testing Group (Foshan) Co., Ltd.

Hotline: 400-840-2288 E-mail: info@waltek.com.cn
[Http://www.waltek.com.cn](http://www.waltek.com.cn)